



EXTENSION MEMO

Page 1 of 1

Product Description(s): Reference Standard 1

Catalog Number(s) / Size(s): C00ADK-2

Lot Number: A0080286

Based on additional testing, the original date of expiration for this Reference Standard has been extended by 60 months. The expiration date of A0080286 has been extended to 28 FEB 2031.

Function	Name	Signature	Date
Quality	Karen Hamilla		06 MAY 2024

This memo supersedes all other statements regarding date of expiration for this product.



RFE MEMO

Page 1 of 1

Product Description(s): Serology Control Pack 1

Catalog Number(s) / Size(s): C4381-1

Lot Number: A00C0825, A00C0826, A00C0827

Based on additional testing, the original date of expiration for this Control Pack has been extended by 60 months. The expiration date of A00C0825, A00C0826, A00C0827 has been extended to 31 MAR 2032

The expiration dates for the following controls have been extended:

Description	Lot Number	Original Expiration Date	New Expiration Date
Serology Control 1.1	A00C0825	31 MAR 2027	31 MAR 2032
Serology Control 1.2	A00C0826	31 MAR 2027	31 MAR 2032
Serology Control 1.3	A00C0827	31 MAR 2027	31 MAR 2032

Function	Name	Signature	Date
Quality	Karen Hamilla		03 MAY 2024

This memo supersedes all other statements regarding date of expiration for this product.



CERTIFICATE OF ANALYSIS

Page 1 of 2

Product Description: V-PLEX COVID-19 Coronavirus Panel 3 Kit

Lot Number: K0082846

Expiration Date: 31 AUG 2027

Catalog Numbers: K15399U-Series, K15400U-Series, K15401U-Series, K15402U-Series

Kit Components:

The kit expiration date provided in the COA is determined by the component with the earliest expiration date. As the kit is supplied with different catalog products, not all configurations include all the components. In such cases, the applicable expiration date is determined by the shortest expiry among the components included in the kit that is received.

Description	Lot Number	Storage Temperature	Expiration Date
Coronavirus Plate 3 96-Well, 10-Spot SECTOR Plate	Z0057636	2-8°C	31 MAY 2028
SULFO-TAG™ Anti-Human IgG Antibody	D00V0038	2-8°C	31 AUG 2027
Reference Standard 1	A0080286	≤ -70°C	28 FEB 2031
Serology Control 1.1	A00C0825	≤ -70°C	31 MAR 2032
Serology Control 1.2	A00C0826	≤ -70°C	31 MAR 2032
Serology Control 1.3	A00C0827	≤ -70°C	31 MAR 2032
ACE2 Calibration Reagent	Not Kit Specific	≤ -70°C	See Label
SULFO-TAG™ Human ACE2 Protein	Not Kit Specific	2-8°C	See Label
SULFO-TAG™ Anti-Human IgM Antibody	Not Kit Specific	2-8°C	See Label
SULFO-TAG™ Anti-Human IgA Antibody	Not Kit Specific	2-8°C	See Label
Diluent 100	Not Kit Specific	2-8°C	See Label
MSD Blocker A Kit	Not Kit Specific	Room Temperature	See Label
MSD GOLD Read Buffer B	Not Kit Specific	Room Temperature	See Label

Plate Uniformity Testing Results:

Parameter	Precision			Uniformity	Signal
	Metric	CV of Intra-plate Averages	Average Intra-plate CV		
SARS-CoV-2 Spike	≤ 18%	≤ 10%	≤ 13%	Pass	1,500 – 1,000,000
HCoV-NL63 Spike	≤ 25%	≤ 15%	≤ 18%	Pass	1,500 – 1,000,000
SARS-CoV-2 Nucleocapsid	≤ 18%	≤ 10%	≤ 13%	Pass	1,500 – 1,000,000
SARS-CoV-1 Spike	≤ 18%	≤ 10%	≤ 13%	Pass	1,500 – 1,000,000
MERS-CoV Spike	≤ 25%	≤ 15%	≤ 18%	Pass	1,500 – 1,000,000
HCoV-HKU1 Spike	≤ 18%	≤ 10%	≤ 13%	Pass	1,500 – 1,000,000
HCoV-OC43 Spike	≤ 18%	≤ 10%	≤ 13%	Pass	1,500 – 1,000,000
HCoV-229E Spike	≤ 25%	≤ 15%	≤ 18%	Pass	1,500 – 1,000,000
SARS-CoV-2 S1 RBD	≤ 18%	≤ 10%	≤ 13%	Pass	1,500 – 1,000,000



CERTIFICATE OF ANALYSIS

Page 2 of 2

Coating Confirmation Testing Results:

Spot	Description	Result
1	SARS-CoV-2 Spike	Pass
2	HCoV-NL63 Spike	Pass
3	SARS-CoV-2 Nucleocapsid	Pass
4	SARS-CoV-1 Spike	Pass
6	MERS-CoV Spike	Pass
7	HCoV-HKU1 Spike	Pass
8	HCoV-OC43 Spike	Pass
9	HCoV-229E Spike	Pass
10	SARS-CoV-2 S1 RBD	Pass

Functional Testing Results:

Sample Type	Calibrator	Controls		Samples	
		Metric	Avg Signal Ratio Test:Reference	Control Recovery	Control Conc. CV
SARS-CoV-2 Spike	50 – 200%	70 – 130%	< 20%	± 25%	0.80 – 1.2
HCoV-NL63 Spike	50 – 200%	70 – 130%	< 20%	± 25%	0.80 – 1.2
SARS-CoV-2 Nucleocapsid	50 – 200%	70 – 130%	< 20%	± 25%	0.80 – 1.2
SARS-CoV-1 Spike	50 – 200%	70 – 130%	< 20%	N/A	N/A
MERS-CoV Spike	50 – 200%	70 – 130%	< 20%	± 25%	0.80 – 1.2
HCoV-HKU1 Spike	50 – 200%	70 – 130%	< 20%	± 25%	0.80 – 1.2
HCoV-OC43 Spike	50 – 200%	70 – 130%	< 20%	± 25%	0.80 – 1.2
HCoV-229E Spike	50 – 200%	70 – 130%	< 20%	± 25%	0.80 – 1.2
SARS-CoV-2 S1 RBD	50 – 200%	70 – 130%	< 20%	± 25%	0.80 – 1.2

Additional Comments:

Functional testing was executed using Coronavirus Plate 3 (Z0057636), SULFO-TAG Anti-Human IgG Antibody (D00V0038), Reference Standard 1 (A0080286), and Serology Controls (A00C0825, A00C0826, A00C0827).

All kit components were manufactured and tested according to MSD documents. The lots listed in the Kit Components table meet MSD's specifications.

Statement:

The above product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
Review/Approval	Karen Hamilla	Quality		29 JAN 2026