

CERTIFICATE OF ANALYSIS

COA - 03677

Product Description: MULTI-ARRAY® 96-well Plate Pack

Catalog #/Size: L15XA-1 / 1 Plate; L15XA-3 / 10 Plates; L15XA-6 / 100 Plates; L15XA-7 / 500 Plates

Lot Number: 21C2ZA

Expiration Date: 31 DEC 2014

Statement: This product was manufactured and tested according to approved MSD documents. This

lot meets MSD's specifications.

BON (Electrical resistivity): < 200 microohms in all circuits

Leak test: No leaks Free Tag 15 000 test

ECL (mean \pm tolerance): 15 000 \pm 3,000 cts Max Average Intraplate CV for any spot: 4% Maximum Intraplate CV for any plate: 7%

Max Interplate CV for any spot: 7%

The above product is intended for Research Use Only. Not for use in Diagnostic Procedures.

	Name	Function	Signature	Date
Preparer	Rolanda Spratley	Biologics/Chemistry	male	18Jan 2013
Review/Approval	Michael Eager	Quality Control	7710	037112013
			7 010	